

**National Institute of Food Technology Entrepreneurship and Management**

Plot 97, Sector 56, HSIIDC Industrial Estate, Kundli, Distt-Sonipat, Haryana  
Email: [niftem.purchasesection@gmail.com](mailto:niftem.purchasesection@gmail.com) Website: [www.niftem.ac.in](http://www.niftem.ac.in)

Date: 17/11/20

**CORRIGENDUM/ADDENDUM**

**Tender No. N/AE/L/2020/6/18 for the Procurement/supply of Gel Imaging System (for AES Lab)**

Reference: 1) Date of Issue of Tender Document: 5/11/20

2) Date of Pre-bid Meeting : 12/11/20

Please refer the pre-bid meeting held on above cited date for the procurement of above equipments.


On the recommendation of User as well as Technical Evaluation Committee, the following corrigendum/addendum is proposed as under:


Clause No.	Present Para/Specification	Proposal Corrigendum/Addendum (to be read)
Point No 4	The Gel Imaging system should offer Smart Tray and adjust imaging parameters and software option accordingly.	The Gel Imaging system should offer Smart Tray / Smart Exposer technology for best capture.
Point No 6	The Gel Imaging System should have a Light - resolution scientific Grade CMOS Camera a light- tight compact darkroom & amp : a Slide -out UV Transilluminator .	The Gel Imaging System should have a Light - resolution scientific grade CMOS Camera and a light- tight compact darkroom & amp with UV/Green Transilluminator.
Point No11	The instrument should offer trans-UV (B) and EPI white as standard Illumination White.	The instrument should offer trans-UV/green and EPI white as standard Illumination.

The said tender may be read amended to this extent. The other terms & condition of tender are remain unchanged.

**Signatures of TEC members**

  
**Dr. Bhim Pratap Singh**  
(Associate professor)

  
**Dr. Neetu Taneja**  
(Assistant professor)

  
**Dr. Tripti Agarwal**  
(Assistant professor)